A Multi-Center Ring Trial of Allergen Exposure Assessment using Fluorescent Multiplex Array Technology

EM King¹, S Filep¹, B Smith¹, P Thorne², R van Ree³, S Arbes⁴, A Calatroni⁴, H Mitchell⁴, MD Chapman¹

¹ Indoor Biotechnologies Inc., Charlottesville, VA, ² Department of Occupational and Environmental Health, University of Iowa, IA, 3 Academic Medical Center, Amsterdam, The Netherlands, 4 Rho, Inc., Chapel Hill, NC

INTRODUCTION

Standardization of allergen exposure measurements is important to ensure reproducibility of results obtained by different laboratories.

We have conducted a Multi-Center Ring Trial with 9 participating centers in the US and Europe to evaluate intra- and inter-laboratory reproducibility of a Multiplex ARray for Indoor Allergens (MARIA™). MARIA ™ enables the simultaneous quantitative detection of eight common indoor allergens of dust mite, cat, dog, mouse, rat and cockroach in a single test. MARIA™ uses xMAP® technology and is ideally suited to large epidemiologic studies and routine environmental assessment of allergen exposure.

METHODS

Study Design

- ➤ 151 dust extracts
- Measured by 9 participating laboratories
- > For 8 indoor allergens using MARIA
- > On 3 separate occasions
- All participants were trained by the coordinating center
- Statistical evaluations were performed by Rho Inc., NC
 - Comparison of over 32 000 allergen measurements

Participating Centers:

U.S.:

Dr. E King

Indoor Biotechnologies Inc., VA (Coordinating center) Dr. D Zeldin NIEHS, Research Triangle Park, NC

Dr. D Schmechel

NIOSH, Morgantown, WV University of Virginia, VA Dr. T Platts-Mills

Dr. R Hamilton

Johns Hopkins University, Baltimore, MD University of Iowa, IA

Dr. P Thorne Dr. D Milton

University of Massachusetts, MA

Europe:

Dr. R van Ree Dr. B Brunekreef Academic Medical Center, Amsterdam, The Netherlands IRAS Div Occ & Envl Health, Utrecht, The Netherlands

RESULTS

Range of Allergen Concentrations: Allergen Concentrations in the 151 dust extracts covered a wide range for all measured allergens, between below detection limit (<LOD) and >130ug/g (Mus m 1: <35ug/g).

INTRA-LABORATORY REPRODUCIBILITY

Results between 3 separate measurements of 151 samples within each study site were evaluated using Intra-class Correlations and Coefficients of Variation (CV%).

- Results within each study site correlated very closely.
- 53% of median CV% within laboratories were within the 5% CV margin and 75% fell within 10% CV.
- The actual level of intra-laboratory reproducibility, however, differed slightly between study sites.


INTER-LABORATORY REPRODUCIBILITY

Results between study sites were evaluated using Intra-class Correlations, Comparison of Overall Means and Coefficients of Variation (CV%).


- Results between labs were highly correlated for all allergens (~0.95).
- Overall means of results were comparable between labs.
- Inter-lab CV% ranged between 19% and 27%.
- Higher variability for Rat n 1 is not significant due to the low number of positive samples

Frequency distribution of intraclass correlations (N = 72)


2007: 119:428-433



Median of CV% between 3 runs for each sample and allergen



Median and range of 8 correlations between each lab and all other labs




CONCLUSIONS

- MARIA results are reproducible both within and between laboratories.
- The actual level of reproducibility is dependent on the operator.
- Precise data interpretation instructions are important.
- MARIA will improve standardization of allergen exposure assessment.
- Continued proficiency testing programs will verify interlaboratory reproducibility.

Comparison of overall means

Co



REFERENCES

ACKNOWLEDGEMENTS

Earle CD, King EM, Tsay A, Pittman K, Saric B, Vailes L, Godbout R, Oliver KG, Chapman MD. High-throughput Fluorescent Multiplex Array for Indoor Allergen Exposure Assessment. J Allergy Clin Immunol

We would like to thank everybody involved in the trial for their hard work! This work was supported in part by a Small Business Innovation Research contract

ES55545 from the National Institute for Environmental Health Sciences.